

PharmLabs San Diego Certificate of Analysis

3421 Hancock St, Second Floor, San Diego, CA 92110 | License: C8-000098-LIC
ISO/IEC 17025:2017 Certification L17-427-1 | Accreditation #85368Sample **Sour Diesel-Cart**

| | | | |
|-------------------|----------------------|--------------|---------------------------------------|
| Sample ID | SD210608-024 (42684) | Matrix | Distillate - NI (Other Cannabis Good) |
| Tested for | The D Buzz | | |
| Sampled - | Received | Jun 08, 2021 | Reported |
| Analyses executed | CAN+ | | |

CAN+ - Cannabinoid Profile Analysis

Analyzed Jun 09, 2021 | Instrument HPLC-VWD | Method SOP-001

| Analyte | LOD mg/g | LOQ mg/g | Result % | Result mg/g |
|---|-------------|-------------|--------------|----------------|
| Cannabidivarin (CBDV) | 0.002 | 0.8 | ND | ND |
| Cannabidiolic Acid (CBDA) | 0.001 | 0.012 | ND | ND |
| Cannabigerol Acid (CBGA) | 0.001 | 0.012 | ND | ND |
| Cannabigerol (CBG) | 0.001 | 0.012 | ND | ND |
| Cannabidiol (CBD) | 0.001 | 0.012 | ND | ND |
| Tetrahydrocannabivarin (THCV) | 0.001 | 0.012 | ND | ND |
| Cannabinol (CBN) | 0.001 | 0.012 | 0.31 | 3.09 |
| Tetrahydrocannabinol (Δ 9-THC) | 0.003 | 0.012 | 0.24 | 2.40 |
| Δ 8-tetrahydrocannabinol (Δ 8-THC) | 0.004 | 0.012 | 83.18 | 831.75 |
| Cannabicyclol (CBL) | 0.002 | 0.006 | ND | ND |
| Cannabichromene (CBC) | 0.002 | 0.005 | ND | ND |
| Tetrahydrocannabinolic Acid (THCA) | 0.001 | 0.012 | ND | ND |
| Total THC (THCa * 0.877 + THC) | | | 0.24 | 2.40 |
| Total CBD (CBDA * 0.877 + CBD) | | | ND | ND |
| Total CBG (CBGa * 0.877 + CBG) | | | ND | ND |
| TOTAL CANNABINOIDS | | | 83.73 | 837.30 |

Sample photography



ND Not Detected
N/A Not Applicable
NT Not Reported
LOD Limit of Detection
LOQ Limit of Quantification
<LOQ Detected
>ULOL Above upper limit of linearity
CFU/g Colony Forming Units per 1 gram
TNTC Too Numerous to Count



Accreditation #85368

Authorized Signature

Dr. Lia Prevedello, Laboratory Director
Wed, 09 Jun 2021 13:22:37 -0700

PharmLabs San Diego | 3421 Hancock St, Second Floor, San Diego, CA 92110 | 619.356.0898 | ISO/IEC 17025:2017 Certification L17-427-1